

CALIBRATION CERTIFICATE | PASS

Test Performed On 23/11/2021 11:56:50

Visual Alarm Pass Audible Alarm Pass

Product Name GasAlertMicroClip	Model Name GasAlertMicroClip	Serial Number KA418-1198858	Firmware Version
Next Cal Due 22/05/2022	Next Bump Due ---	Unit Programmed ---	
Location ---	Assigned User ---		

Test Station

Product Name MicroDock II	Model Name MicroDock II	Serial Number Z206-003705	Firmware Version V0.0.0
------------------------------	----------------------------	------------------------------	----------------------------

Inlet Information

Inlet Number 1	Inlet Notes	Gases Purge	Concentration 20.9	Inlet Used Yes
Inlet Number 2	Inlet Notes 9125:4713	Gases 4 Gas Mixture Equivalent	Concentration 25.0	Inlet Used Yes
Inlet Number 3	Inlet Notes	Gases Purge	Concentration 20.9	Inlet Used No
Inlet Number 4	Inlet Notes	Gases Purge	Concentration 25.0	Inlet Used No

Sensor Information

Sensor	Results	Next Cal Due	Due In
H2S	Pass	22/05/2022	180 Days
CO	Pass	22/05/2022	180 Days
LEL	Pass	22/05/2022	180 Days
O2	Pass	22/05/2022	180 Days

Zero Calibration | PASS

Sensor	Gas Result
H2S (ppm)	Pass
CO (ppm)	Pass
LEL (%LEL)	Pass
O2 (%VOL)	Pass

Set Points

Sensor	Low	High	STEL	TWA
H2S (ppm)	5.0	10.0	10.0	5.0
CO (ppm)	20.0	100.0	100.0	20.0
LEL (%LEL)	10.0	20.0	---	---
O2 (%VOL)	19.5	23.5	---	---

Certificate Generated On 23/11/2021 12:35:28